

The Transmission Electron Microscope

Karen M. Davies

Principal Beamline Scientist – eBIC,
Diamond Light Source Ltd, UK

Wellcome Trust Graduate Training Program - CryoEM

Outline

Basic image formation – eye vs. microscope

Hardware of a transmission electron microscope:

Vacuum system

Electron source

Types of lenses

Apertures

How the microscope works:

Condenser system

Objective lens

Projection system

Aberrations and correctors

Detection system

Suggested Reading


Basic Image Formation

Illumination source

Object

Lens

Detector




Basic Image Formation

Illumination source


Object

Lens


Detector




Microscope



Light vs. Electron Microscope




Light




Electron

The Vacuum System




Vacuum system



Electron

The Vacuum System



Vacuum system

Why operate in vacuum?

- Electrons interact well with matter
- Mean-free-path length: 20 cm in air
2 km in vacuum

How is vacuum achieved?

- Rotary pump (PVP) – ATM to rough vac.
- Oil diffusion pump (ODP) / Scroll Pump - low vac
- Turbo pump – high vacuum
- Ion getter pump (IGP) – ultra high vacuum
- Cryo-pump/trap – high vacuum

How is vacuum monitored?

- Pirani gauge (Pir) – ATM to low vac
- Penning gauge (Pen) – high vac
- Current readout (IGP) – ultra high vac

$$1\text{Pa} = 0.01\text{ mbar}$$


- Room Pressure: 10^5 Pa
- Rough Vacuum: 100-0.1 Pa
- Low Vacuum: $0.1 - 10^{-4}$
- High Vacuum: $10^{-4} - 10^{-7}$ Pa
- Ultra High Vacuum: 10^{-7} Pa and below

Vacuum System


Atmosphere

Low vacuum


Low vacuum



Rotary pump



Scroll Pump




Turbo Pump

Increasing Vacuum


High vacuum

high vacuum




Oil Diffusion Pump

Low vacuum




Ion Getter Pump




Cryo-pump

Electron Source

Thermionic



Field Emission




Gun quality:

Temporal coherence – Wavelength spread
Spatial coherence – Angular spread

	Units	Tungsten	LaB ₆	Schottky FEG	X-FEG	Cold FEG
Work function, Φ	eV	4.5	2.4	3.0		4.5
Richardson's constant	A/m ² K ²	6×10^9	4×10^9			
Operating temperature	K	2700	1700	1700		300
Current density (at 100 kV)	A/m ²	5	10^2	10^5		10^6
Crossover size	nm	$> 10^5$	10^4	15		3
Brightness (at 100 kV)	A/m ² sr	10^{10}	5×10^{11}	5×10^{12}		10^{13}
Energy spread (at 100 kV)	eV	3	1.5	0.7		0.3
Emission current stability	%/hr	<1	<1	<1		5
Vacuum	Pa	10^{-2}	10^{-4}	10^{-6}		10^{-9}
Lifetime	hr	100	1000	>5000		>5000
Cost of tip		£80	£800	£8000		£8000
Time to replace		1-2 days	1-2 days	5-8 days		5-8 days


Filaments

Tungsten (W)






Bozzola and Russell, Fig. 6.22

LaB₆




FEG (W)




Thermionic

Field Emission

The Lens System – Types of Lenses



The Lens System – Types of Lenses



Condensing lenses

Condenser system
Focuses electrons on specimen
(controls illumination cone shape/intensity)

Objective system
Main Magnification Lens
(magnifies sample, 80/100x)




Projector system
Weaker magnification lenses
(magnifies sample, 2-10x)

Apertures
Reduces beam intensity
Parallelizes beam
Increases amplitude contrast



Deflector coils Corrects beam location/angle (tilt/shift)

Quadrupoles Corrects beam shape (stigmator)

The Imaging (Condensing) Lens



The Imaging (Condensing) Lens




Magnetic lenses are poor quality and have severe aberrations:

Spherical aberration (position)

Chromatic aberration (wavelength)


Coma (angle of entry)

Stigmatism – defects in magnetic field symmetry




Aberrations from Condensing Lenses

Spherical C_s




- Lenses are strongest at edge
- Thus off-axis electrons bent more than on-axis
- Different focus points

Chromatic C_c



- Electrons of different wavelength focused by different amount
- Focus plane different for different wavelengths

Coma



- Beam enters lens at angle
- C_s causes rays to bend depending on location
- Point source becomes comet shaped




Quadrupoles = Stigmators

Stigmatism is caused by:

- 1) Magnetic lens defects
- 2) Contamination in pole gap




Condenser stigmators
correct beam shape
(circle vs. oval)

Objective stigmators
correct image




Objective stigmatism

Shift/Tilt Coils




1. Get the beam down the scope
2. Make sure beam is parallel/aligned to optical axis (bright field imaging)
3. Low dose imaging
4. Automatic procedures (e.g. eucentric height/focus)


Use of Deflection Coils in Low Dose Imaging



Apertures




Condenser aperture
Objective aperture
Selective area aperture




Controls beam intensity, parallity
Amplitude contrast
Diffraction imaging/dark field

What is happening when I operate the microscope?



Condenser System



Gun tilt/shift: sets up beam to enter condenser system on optical axis

C1 = Spot size, controls beam size and quality


Spot 9: Strongest lens setting
Highest crossover
Dimmest beam and smallest focused beam
Most coherent little spatial divergence
Only most parallel electrons reach specimen

Spot 1: Weakest lens setting
Lowest crossover
Brightest & largest focus beam diameter
Least coherent
Greatest spatial divergence
Majority of electrons reach specimen


C2 aperture = limits amount of electrons reaching sample

C2 lens = Intensity knob. Controls diameter of beam reaching sample → (Dose/beam intensity)

Note on C2 Lens and Dose



$$D_A = \frac{D_T}{A} = \frac{D_T}{\pi r^2}$$




Dose in an area increase 4x when
you half the radius.

**USE THIS INTENSITY CONTROL
KNOB WISELY!**

DA = Dose per unit area
DT = Total dose In beam
A = Area
r = Beam Radius


C3 Lens (Krios)



Parallel illumination of specimen reduces aberrations




C3 lens provides parallel illumination but only at certain C2 values

C3 Lens (Krios)





High tension:	300 kV	Unavail.	X:	-853.87 μm	C2 Lens:	40.687
Screen current:	0.000 nA	Illuminated area:	Y:	-290.48 μm	C3 Lens:	48.836
Defocus:	909.97 nm	Spot size:	Z:	-34.50 μm	Obj Lens:	81.7245
Focus step:	5	5	A:	-0.01 deg	Cooling BM-Falcon:	Stat
			B:	1.01 deg	Cooling BM-Ceta:	Unknov
			C1 Lens:	31.603 %		


Objective System – Mag and Specimen



Objective System – Mag and Specimen




- Cryoblades cooled to LN₂ temp by braids
- Cryoblades colder than sample → cryotrap
- Keeps sample clean and cold by absorbing contaminant from sample/column
- Cryoblades needs to be warmed up regularly to remove contaminants → cryocycle
- Cryocycle: turns off IGP, pumps specimen chamber with Turbo



Gap between pole pieces is spherical aberration (Cs). Smaller gap = better.


Objective System - Eucentric Height



Excellent!

Tilt axis aligns on optical axis and eucentric focus


Specimen does not move when tilted



Bad

Specimen moves when tilted

Change specimen Z-height







Very Bad

Specimen moves when tilted

Change specimen Z-height
Software or engineer required to align tilt axis with optical axis

Objective System - Focus



Modulating objective lens current changes clarity of image (focus)

If crossover above image plane - overfocused

If crossover below image plane - underfocused



Biological samples contain light atoms so minimal phase shifts occurs between scattered and unscattered rays

→ At focus, very little contrast

To see sample, either image underfocus or use phase plate


Imaging underfocus = CTF issues

Objective System - Phase Plates




5 μ m defocus

At focus with VPP



Projection System



Condenser system

Objective system

Projector system

- Bunch of low strength lenses to increase magnification
- Tend to be ignored as aberrations outweighed by objective lens

Aberration Correctors

Magnetic lenses are poor quality and have severe aberrations:


Spherical aberration (position)

Chromatic aberration (wavelength)

Coma (angle of entry)

Stigmatism – defects in magnetic field symmetry

Spherical Aberration (C_s) – C_s Corrector



- Lenses are strongest at edge
- Thus off-axis electrons bent more than on-axis
- Different focus points





FIGURE 6.12. Ray diagrams showing how the two different commercial systems use (A) multiple quadrupole (Q) and octupole (O) lenses (Nion) or (B) hexapole and other transfer lenses (CEOS) to correct for C_s .

To reduce C_s :

- Increase lens strength
- Decrease pole gap
- Install C_s corrector (important for imaging $< 0.5 \text{ \AA}$)

Chromatic Aberration (Cc)



Electrons of different wavelengths are focused at different point

Cause of wavelength variation:

Electron source

Interaction with sample

(especially thick samples)

Correct using:


FEG instead of Tungsten

Monochromator after gun


Energy filter after sample imaging

(Important for thick samples)


Energy Filter



Unfiltered



20eV Energy Filter



- Elaborate mass spec
- Select for specific wavelength
- Zero-loss imaging (elastically scattered waves)
- EELS – Selects wavelength of inelastically scattered rays – chemical composition

Coma

- Beam enters lens at angle
- Cs causes rays to bend depending on location
- Point source becomes comet shaped
- **To correct: Apply +/- beam tilt**
- FFT of opposite beam tilts should be identical






Image of point
looks like a comet




(Important to obtain resolutions beyond 5 Å)

Detection System

Fluorescent Screen



Charge coupled device (CCD)




Film + Scanner




Direct Electron Detectors (DEDs)

(e.g. K3/Falcon)



Camera Quality Measurement

Modular transfer function (MTF)




How fast does intensity change at sharp edge?

Fast change, great camera, able to capture high resolution data

Slow change, poor camera, resolution limiting

Detection Quantum Efficiency (DQE)



$$DQE = \frac{SN_{out}^2}{SN_{in}^2}$$

DQE = 1, excellent camera, no loss of signal
Nyquist = 1/2px





Figure 3. Computation of the Modulation Transfer Function using the knife-edge target.


Integrated vs Counted (super-resolution) Mode

Pixel




Incoming Electron

Electron hits pixel array and causes charge spread




Integration mode

Integrating mode reads out accumulated charge in each pixel



Counting mode

Assumes $1e^-$ in, $1e^-$ out. Calculates most probably location of electron hit.




Super-resolution

Subdivides pixel to give more accurate location of electron hit.

Nyquist: maximum resolution of detector ($2 \times$ px size)
represented by sine wave where 1px white, next black (smallest sampling frequency)

Suggested Reading



http://myscope.training/#/TEMlevel_2_4

Flash of TEM:

<http://www.doitpoms.ac.uk/tlplib/tem/illumination.php>

<http://cryo-em-course.caltech.edu/overview>

Youtube

[1.7. Electron Optics — Introduction to Transmisison Electron Microscopy \(gduscher.github.io\)](https://github.com/gduscher)